

DAQP-ACC-A Module

TECHNICAL REFERENCE MANUAL

WELCOME TO THE WORLD OF DEWETRON!


Congratulations on your new device! It will supply you with accurate, complete and reproducible measurement results for your decision making.

Look forward to the easy handling and the flexible and modular use of your DEWETRON product and draw upon more than 30 years of DEWETRON expertise in measurement engineering.

IS09001



THE MEASURABLE DIFFERENCE.



Notice

The information contained in this document is subject to change without notice.

DEWETRON GmbH (DEWETRON) shall not be liable for any errors contained in this document. DEWETRON MAKES NO WARRANTIES OF ANY KIND WITH REGARD TO THIS DOCUMENT, WHETHER EXPRESS OR IMPLIED. DEWETRON SPECIFICALLY DISCLAIMS THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. DEWETRON shall not be liable for any direct, indirect, special, incidental, or consequential damages, whether based on contract, tort, or any other legal theory, in connection with the furnishing of this document or the use of the information in this document.

Warranty Information:

A copy of the specific warranty terms applicable to your DEWETRON product and replacement parts can be obtained from your local sales and service office.

Restricted Rights Legend:

Use austrian law for duplication or disclosure.

DEWETRON GmbH Parkring 4 8074 Grambach Austria

Printing History:

Please refer to the page bottom for printing version.

Copyright © DEWETRON GmbH.

This document contains information which is protected by copyright. All rights are reserved. Reproduction, adaptation, or translation without prior written permission is prohibited, except as allowed under the copyright laws.

All trademarks and registered trademarks are acknowledged to be the property of their owners.

Notice

Safety symbols in the manual



Indicates hazardous voltages.

WARNING Calls attention to a procedure, practice, or condition that could cause bodily injury or death.

CAUTION Calls attention to a procedure, practice, or condition that could possibly cause damage to equipment or permanent loss of data.

WARNINGS

The following general safety precautions must be observed during all phases of operation, service, and repair of this product. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the product. DEWETRON Elektronische Messgeraete Ges.m.b.H. assumes no liability for the customer's failure to comply with these requirements.

All accessories shown in this document are available as option and will not be shipped as standard parts.

Safety Instructions

Safety instructions for DEWETRON amplifiers

- The DEWETRON data acquisition systems and amplifiers may only be installed by experts.
- Read your manual carefully before operating.
- Observe local laws when using the amplifiers.
- Ground the equipment: For Safety Class 1 equipment (equipment having a protective earth terminal), a non interruptible safety earth ground must be provided from the mains power source to the product input wiring terminals or supplied power cable.
- DO NOT operate the product in an explosive atmosphere or in the presence of flammable gases or fumes.
- DO NOT operate damaged equipment: Whenever it is possible that the safety protection features built into this product have been impaired, either through physical damage, excessive moisture, or any other reason, REMOVE POWER and do not use the product until safe operation can be verified by service-trained personnel. If necessary, return the product to a DEWETRON sales and service office for service and repair to ensure that safety features are maintained.
- Keep away from live circuits: Operating personnel must not remove equipment covers or shields. Procedures involving the removal of covers or shields are for use by service-trained personnel only. Under certain conditions, dangerous voltages may exist even with the equipment switched off. To avoid dangerous electrical shock, DO NOT perform procedures involving cover or shield removal unless you are qualified to do so.
- No modifications are allowed at the amplifiers.
- DO NOT service or adjust alone. Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present.
- DO NOT substitute parts or modify equipment: Because of the danger of introducing additional hazards, do not install substitute parts or perform any unauthorized modification to the product. Return the product to a DEWETRON sales and service office for service and repair to ensure that safety features are maintained.
- DO NOT touch internal wiring!
- DO NOT use higher supply voltage than specified!
- Use only original plugs and cables for harnessing.
- Safety of the operator and the unit depend on following these rules.

Support

For any support please contact your local distributor first or DEWETRON directly.

For Asia and Europe, please contact:

DEWETRON GmbH Parkring 4 8074 Grambach AUSTRIA

Tel.: +43 316 3070 Fax: +43 316 307090

Email: support@dewetron.com Web: http://www.dewetron.com

The telephone hotline is available Monday to Friday between 08:00 and 12:00 CET (GMT -1:00) and Monday to Thursday between 13:00 and 17:00 CET.

For the Americas, please contact:

DEWETRON, Inc. 2850 South County Trail, Unit 1 East Greenwich, RI 02818 U.S.A.

Tel.: +1 401 284 3750 Toll-free: +1 866 598 3393 Fax: +1 401 284 3755

Email: us.support@dewetron.com Web: http://www.dewetron.us

The telephone hotline is available Monday to Friday between 08:00 and 17:00 GST (GMT +5:00)

General Module Information

Calibration information

All DEWETRON modules are calibrated at 25 °C after a warmup time of 30 minutes and meet their specifications when leaving the factory.

The time interval for recalibration depends on environmental conditions. Typically, the calibration should be checked once a year.

Calibration certificates are available from DEWETRON as an option. DEWETRON offers two types:

- ISO traceable DEWETRON certificate
- Calibration certificate according to ÖKD (equivalent to DKD)

This manual contains no calibration information. For self calibration, there is a separate calibration kit for the DAQ series modules available. The CAL-KIT contains the required cables, software and instructions.

General module specifications

Module dimensions: 20 x 65 x 105 mm (0.79 x 2.56 x 4.13 in.)

(W x H x D without front cover and connectors)

Frontcover: 20 x 87 x 2 mm (0.79 x 3.43 x 0.08 in.)

(W x H x D without connector)

Environmental:

Temp. range storage: -30 °C to +85 °C (-22 °F to 185 °F) Temp. range operating: -5 °C to +60 °C (23 °F to 140 °F)

Relative humidity

(MIL202): 0 to 95 % at 60 °C, non-condensing (unless otherwise noticed)

RFI susceptibility: ±0.5 % span error at 400 MHz, 5 W, 3 m

All specifications within this manual are valid at 25 °C!

All modules are produced according ISO9001 and ISO14001.

Module connectors

Frontpanel connector: Accessable to the user. The connector type and pin assignment varies

from module to module. Detailed pin assignment of each module is

shown in the appropriate module description.

Rear connector: 9-pin male SUB-D, interface to the DEWE-System, not accessable to the user.



9-pin male SUB-D connector

HSI/DAQx and PAD module rear view


Interface pin assignment:

- 1 Module output (±5 V)
- 2 RS-485 (A)
- 3 RS-485 (B)
- 4 GND
- 5 +9 V power supply
- 6 +12 V power / sensor supply
- 7 Module input (from D/A converter of the A/D board)¹¹)
- 8 reserved
- 9 -9 V power supply
- 1) Triggerout at DAQP-FREQ-A

General Module Information

RS-232/485 interface

HSI/DAQP modules can be configured via RS-485 interface, PAD modules require this interface for all data transfers.



For all DEWETRON systems, an internal RS-232/485 converter is available

(standard with DEWE-800, -2000, -2500, -3000, -4000, -5000 series systems). This converter allows communication with HSI/DAQP and PAD modules.

To communicate with the modules, the RS-232 interface has to be set to the following parameters:

baud rate: 9600 bps

data bits:

parity: no parity

stop bits:

handshake: not required

HSI/DAQP module configuration

1. Push button selection

All ranges and filters can be selected directly by pressing the push buttons on the module. Approx. 15 sec. after changing range and / or filter, the range and filter information is stored in an EEPROM. This procedure increases the livetime of the EEPROM.

The current input range setting is shown all the time by LED. To change the range just press *RANGE* button a few times until the required range is displayed.

To see the current filter setting just press the *FILTER* button once. The corresponding LED is flashing for approx. 3 seconds. Within this time, the filter can be selected by pressing the *FILTER* button again. Approx. 3 seconds after the last key activity, the information will be stored, the LED stops flashing and shows the input range again.

CAUTION: Power loss during this time leaves the module in the former settings.

2. RS-232/485 programming

All ranges and filters can also be selected via RS-232/485 interface. All new DEWE-800, -2000, -2500, -3000, -4000, -5000 series systems are prepared as a standard to work with HSI/DAQP modules.

The easiest way to change the configuration is to use the DEWEConfig software, which comes as a standard with the DEWETRON data acquisition system.

Detailed information about HSI/DAQP modules programming for customer applications is available in the *DEWE-Modules Programmers Reference Manual*.

CAUTION: All range and filter changes which are done via RS-232/485 interface are not stored in the EEPROM of the HSI/DAQP modules! You have to store this information in a separat initialisation file to keep settings information for next system start!

PAD module communication

All PAD modules are only working through the RS-232/485 interface. All new DEWE-800, -2000, -2500, -3000, -4000, -5000 series systems are prepared as a standard to work with PAD modules. The easiest way to change the configuration is to use the DEWEConfig software, which comes as a standard with the DEWETRON data acquisition system.

Detailed information about PAD modules programming for customer applications is available in the *DEWE-Modules Programmers Reference Manual*.

DAQP-ACC-A

High dynamic signal amplifier

■ Supported sensors: IEPE® sensors

■ Input sensitivity: ±5 V, ±1.66 V, ±500 mV, ±166 mV, ±50 mV

(Gain 1, 3, 10, 30 and 100), push button or software selection

■ Bandwidth: 300 kHz

Sensor connection: BNC connector



Module specifications

	DAQP-ACC	-A				
Ranges	±5 V, ±1.66 V, ±500 mV, ±166 mV, ±50 mV					
Gain	1, 3, 10, 30, 100					
Range / gain selection	Push button or software					
Gain error	0.5 %					
Sensor types	IEPE® sensors only					
Sensor excitation	4 or 8 mA (software selection), 10 %, up to 28 $\rm V_{\rm DC}$					
Input impedance	5 or 7 MOhm (depending on time constant), in parallel with 1.2 nF					
Input voltage range Voltage < 4 V Voltage > 19 V	4 to 19 V ,Shortcut' detection ,No sensor' detection					
Input protection IN+ IN- (shield)	max10 to 28 V max. 20 mA					
Bandwidth (-3 dB)	From selected highpass filter to 300 kHz (+2 to -5 dB @ fg)					
Filters (highpass) 0.5 Hz filter 5 Hz filter Filters (lowpass)	0.5 Hz and 5 Hz (software selection) 0.32 s time constant 0.032 s time constant 1 kHz, 10 kHz, 300 kHz					
	other filter steps available as an option on request					
Filter selection	Push button or software					
Filter characteristics up to 100 kHz 300 kHz	Butterworth 100 dB / decade (30 dB / octave) 80 dB / decade (24 dB / octave)					
Typical SFDR and SNR	300 kHz bandwidth 100 kHz bandwidth			10 kHz bandwidth		
5000 mV 500 mV 50 mV	SFDR 92 dB 88 dB 71 dB	SNR 90 dB 85 dB 68 dB	SFDR 100 dB 100 dB 90 dB	SNR 91 dB 89 dB 73 dB	SFDR 103 dB 100 dB 82 dB	SNR 93 dB 92 dB 80 dB
Output voltage	±5 V					
Output resistance	< 10 Ohm					
Output current	Max. 5 mA					
Output protection	Continuous short to ground					
RS-485 interface	Yes					
Power supply voltage	±9 V _{DC} (±10 %)					
Power consumption	Typical 0.8 to 1.0 W (depending on sensor)					

V

DAQP-ACC-A

LED state

The DAQP-ACC-A series module has a set of 6 LEDs showing the current input range (constant active) and filter range (flashing) setting. The ERR LED will displays shortcut and no sensor connected.

Input gain and filter selection

The DAQP-ACC-A series module has two push buttons.

■ Gain button: Push the *GAIN* button several times until the LED displays the desired input

range.

■ Filter button: Push the *FILTER* button once - the LEDs will flash for approx. 3 seconds and


display the current filter setting.

Push the *FILTER* button within the three seconds several times until the

flashing LED displays the desired filter range.


Block diagram

The base block diagram of the DAQP-ACC-A gives an idea of the internal structure.



Sensor connection

IEPE® sensor connection



CE-Certificate of conformity



Manufacturer:

Address:

DEWETRON GmbH

Parkring 4 8074 Grambach, Austria

Tel.: +43 316 3070 0 Fax: +43 316 3070 90 e-mail: sales@dewetron.com http://www.dewetron.com

Name of product:

DEWE-MODULES

Kind of product: Signal conditioning modules

The product meets the regulations of the following EC-directives:

73/23/EEC

"Directive on the approximation of the laws of the Member States relating to electrical equipment designed for use within certain voltage limits amended by the directive 93/68/ EEC"

89/336/EEC

"Directive on the approximation of the laws of the Member States relating to electromagnetic compatibility amended by the directives 91/263/EEC, 92/31/EEC, 93/68/EEC and 93/97/EEC

The accordance is proved by the observance of the following standards:

L V	Safety	IEC/EN 61010-1:1992/93 IEC/EN 61010-2-031 IEC 1010-	IEC 61010-1:1992/300 V CATIII Pol. D. 2 2-031
E	Emissions	EN 61000-6-4	EN 55011 Class B
C	Immunity	EN 61000-6-2	Group standard

Graz, April 28, 2010

Place / Date of the CE-marking

Dipl.-Ing. Roland Jeutter / Managing director

▼ Notes