



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

DEWETRON Inc.
2850 County Trail
East Greenwich, RI 02818

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 02 June 2027

Certificate Number: AC-3212



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

DEWETRON Inc.
2850 South County Trail
East Greenwich, RI 02818
401-284-3750

CALIBRATION

Valid to: **June 2, 2027**

Certificate Number: **AC-3212**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source & Measure	Up to 100 mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	24 $\mu\text{V/V} + 0.3 \mu\text{V}$ 6.8 $\mu\text{V/V} + 0.3 \mu\text{V}$ 7.9 $\mu\text{V/V} + 0.5 \mu\text{V}$ 8.2 $\mu\text{V/V} + 30 \mu\text{V}$ 13 $\mu\text{V/V} + 0.1 \text{ mV}$	Comparison to Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
DC Current – Source & Measure	Up to 1 mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A (1 to 11) A (11 to 20) A	16 $\mu\text{A/A} + 5 \text{ nA}$ 27 $\mu\text{A/A} + 50 \text{ nA}$ 36 $\mu\text{A/A} + 0.5 \mu\text{A}$ 110 $\mu\text{A/A} + 10 \mu\text{A}$ 400 $\mu\text{A/A} + 0.39 \text{ mA}$ 770 $\mu\text{A/A} + 0.58 \text{ mA}$	Comparison to Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
DC Current – Source & Measure	(0.2 to 1) A (1 to 10) A (10 to 20) A	0.003 2 % of reading + 2.3 μA 0.004 5 % of reading - 11 μA 0.007 6 % of reading - 320 μA	Comparison to Agilent 3458A Multimeter with Fluke A40 Shunt
DC Resistance – Measure	Up to 10 Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω (0.1 to 1) M Ω	16 $\mu\Omega/\Omega + 50 \mu\Omega$ 8.0 $\mu\Omega/\Omega + 0.5 \text{ m}\Omega$ 6.9 $\mu\Omega/\Omega + 0.5 \Omega$ 7.4 $\mu\Omega/\Omega + 5 \Omega$ 7.2 $\mu\Omega/\Omega + 50 \Omega$ 12 $\mu\Omega/\Omega + 2 \Omega$	Comparison to Agilent 3458A Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance - Source	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω (110 to 330) M Ω (330 to 1 100) M Ω	55 $\mu\Omega/\Omega$ + 1.0 m Ω 47 $\mu\Omega/\Omega$ + 1.5 m Ω 25 $\mu\Omega/\Omega$ + 1.4 m Ω 22 $\mu\Omega/\Omega$ + 2.0 m Ω 29 $\mu\Omega/\Omega$ + 2.0 m Ω 22 $\mu\Omega/\Omega$ + 20 m Ω 25 $\mu\Omega/\Omega$ + 20 m Ω 37 $\mu\Omega/\Omega$ + 0.2 Ω 25 $\mu\Omega/\Omega$ + 0.2 Ω 27 $\mu\Omega/\Omega$ + 2 Ω 29 $\mu\Omega/\Omega$ + 2 Ω 73 $\mu\Omega/\Omega$ + 30 Ω 0.013 % of reading + 50 Ω 0.038 % of reading + 2.5 k Ω 0.061 % of reading + 3 k Ω 0.4 % of reading + 0.1 M Ω 1.2 % of reading + 0.5 M Ω	Comparison to Fluke 5522A Multi Product Calibrator
AC Voltage – Source & Measure	Up to 100 mV (20 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (0.1 to 1) V (20 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (1 to 10) V (20 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (10 to 100) V (20 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (100 to 1 000) V (20 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz	0.006 % of reading + 4 μ V 0.015 % of reading + 2 μ V 0.068 % of reading + 2 μ V 0.006 % of reading + 40 μ V 0.014 % of reading + 20 μ V 0.074 % of reading + 20 μ V 0.006 % of reading + 0.4 mV 0.014 % of reading + 0.2 mV 0.084 % of reading + 0.2 mV 0.016 % of reading + 4 mV 0.017 % of reading + 2 mV 0.11 % of reading + 2 mV 0.043 % of reading + 40 mV 0.059 % of reading + 20 mV 0.3 % of reading + 20 mV	Comparison to Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source & Measure	(1 to 10) V (16 to 850) Hz (10 to 100) V (16 to 850) Hz (100 to 1 000) V (16 to 850) Hz	0.005 2 % of reading + 0.17 mV 0.005 % of reading + 1.3 mV 0.005 % of reading + 9 mV	Comparison to Fluke 6105A Electrical Power Standard
AC Current – Source & Measure	Up to 100 μ A 45 Hz to 5 kHz (0.1 to 1) mA 45 Hz to 5 kHz (1 to 10) mA 45 Hz to 5 kHz (10 to 100) mA 45 Hz to 5 kHz (0.1 to 1) A 45 Hz to 5 kHz (1 to 3) A 45 Hz to 5 kHz (3 to 11) A 45 Hz to 5 kHz (11 to 20) A 45 Hz to 5 kHz	0.06 % of reading + 30 nA 0.06 % of reading + 0.2 μ A 0.06 % of reading + 2 μ A 0.06 % of reading + 20 μ A 0.1 % of reading + 0.2 mA 0.48 % of reading + 0.78 mA 2.4 % of reading + 1.6 mA 2.3 % of reading + 3.9 mA	Comparison to Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
AC Current – Source & Measure	(0.01 to 0.25) A (16 to 850) Hz (0.25 to 1) A (16 to 850) Hz (1 to 2) A (16 to 850) Hz (2 to 5) A (16 to 850) Hz (5 to 10) A (16 to 850) Hz (10 to 21) A (16 to 850) Hz	0.006 % of reading + 5 μ A 0.006 % of reading + 20 μ A 0.006 % of reading + 40 μ A 0.006 4 % of reading + 100 μ A 0.006 5 % of reading + 200 μ A 0.007 1 % of reading + 400 μ A	Comparison to Fluke 6105A Electrical Power Standard

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouples - Measure/Source	Type J (-200 to -100) °C (-100 to 760) °C (760 to 1 200) °C Type K (-200 to -100) °C (-100 to 1 000) °C (1 000 to 1 370) °C Type T (-250 to -150) °C (-150 to 400) °C	0.3 °C 0.2 °C 0.23 °C 0.35 °C 0.25 °C 0.35 °C 0.64 °C 0.17 °C	Comparison to Fluke 5522A Multi Product Calibrator
Electrical Simulation of RTDs - Measure/Source	Pt385, 100 Ohm (-200 to 630) °C Pt3926, 100 Ohm (-200 to 630) °C Pt385, 200 Ohm (-200 to 630) °C Pt385, 500 Ohm (-200 to 600) °C Pt385, 1 000 Ohm (-200 to 600) °C	0.11 °C 0.11 °C 0.14 °C 0.094 °C 0.063 °C	Comparison to Fluke 5522A Multi Product Calibrator
DC Power Source	10.9 µW to 1W	190 µW/W	Comparison to Fluke 5522A Multi Product Calibrator with Agilent 3458A Multimeter
DC Power Source	(0.15 to 180) W (180 to 720) W (720 to 2 016) W Up to 1 800 W (1 800 to 7 200) W (7 200 to 20 160) W	260 µW/W 260 µW/W 260 µW/W 260 µW/W 260 µW/W 260 µW/W	Comparison to Fluke 6105A Electrical Power Standard
AC Power Source Power Factor PF = 1	16 Hz to 850 Hz (0.15 to 180) W (180 to 720) W (720 to 2 016) W 16 Hz to 850 Hz (1.5 to 1 800) W (1 800 to 7 200) W (7 200 to 20 160) W	120 µW/W 120 µW/W 110 µW/W 130 µW/W 120 µW/W 120 µW/W	Comparison to Fluke 6105A Electrical Power Standard

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power Source			
Power Factor PF = (< 1 to ≥ 0.9)	16 Hz to 450 Hz (0.15 to 20 160) W	200 μ W/W	Comparison to Fluke 6105A Electrical Power Standard
PF = (< 0.9 to ≥ 0.5)	16 Hz to 180 Hz (0.15 to 20 160) W	580 μ W/W	
PF = (< 0.5 to ≥ 0.1)	45 Hz to 65 Hz (0.15 to 20 160) W	3 200 μ W/W	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-3212.



Jason Stine, Vice President